

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To

The Quality of Industrial Products Testing Laboratories.
General Organization for Export & Import Control (GOEIC).
(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-): Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Steel Reinforcing Bars	Mechanical test : Yield strength Tensile Strength Elongation up to 100 Ton	ASTM A370-23 ASTM A29/A29M-20 SHIMADZU Universal Tensile testing machine 100 ton Model: UH-100A S.N: 30096610 Digital Outside Micrometer Model: N/A S/N:67337341

Ayman

Marie El Desouki



Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To

The Quality of Industrial Products Testing Laboratories,
General Organization for Export & Import Control (GOEIC),
(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-): Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Stainless steel Plate, Sheet and Strip (20 ton)	Mechanical test : Yield strength Tensile Strength Elongation Up to 20 Ton	ASTM A370-23 ASTM A240/240M-22	Universal Tensile testing machine 20 ton Model: 5585 S.N: 5585J1714 Digital Outside Micrometer Model: N/A S/N:67337341 Steel Rular Code:ES1830

Amr
Hani El Desouki

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025

Issued To

The Quality of Industrial Products Testing Laboratories.
General Organization for Export & Import Control (GOEIC).
(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-): Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Ceramic Tiles	Determination of Modulus of Rupture and Breaking Strength	ES: 293-10/2005 ISO 10545-4:2019	Breaking strength equipment S.N. : TR 749 Vernier caliper Code : 01 Steel ruler Code : E-01 Oven Code: 1-2017
	Determination of Resistance to Deep Abrasion for Unglazed Tiles	ES: 293-2/2005 ISO 10545-6:2010	Abrasion apparatus S.N. : A117 Vernier caliper Code : 01

Ayman
Hanie El Desouki

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To

The Quality of Industrial Products Testing Laboratories.
General Organization for Export & Import Control (GOEIC).
(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-): Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Low -Alloy Steel	Chemical analysis using atomic emission vacuum spectrometric analysis. C 0.05% to 1.05% Si 0.005% to 1.0% S 0.01 % to 0.050% Cr 0.05 % to 4.8% Mo 0.10% to 1.2% Mn 0.10% to 1.75%	ASTM E415/2021 Optical emission Spectro Lab Manufacturer: SPECTRO Analytical Instruments, GmbH Model : LAB LAVM10 S.N: 10000895

A. El-Dewany
Hanie El Dewany

Ministry of
Trade and Industry
Egyptian Accreditation Council
EGAC



وزارة التجارة والصناعة
الجلس الوطني للاعتماد
إيجاك

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025

Issued To

The Quality of Industrial Products Testing Laboratories.
General Organization for Export & Import Control (GOEIC).

(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-):

Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Water Ball Valves	1- Proof Pressure Test 2- Internal Leakage Test Pressure rating up to 120 bar Nominal size up to 4 inch	EN 12266- 1/2012 EN13828/2003 Ball Valve Testing Machine Model: OTE 000500-S-2000 S.N.: 00500 Inlet Pressure Transducer S.N.: 2898548 Outlet Pressure Transducer S.N.: 2898516

Aghmar

Marie El Desouki

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt
Tel.: (202) 25275220/5/6/7
Fax: (202) 25275224

F4W14TCL
1 / Dec 2018

Page 5 of 6

Industrial Investment Map: <http://invegypt.com>: مصر الصفحة الرئيسية خريطة الاستثمار الصناعي في مصر

كورنيش المادي - برج رياض المادي ١ - القاهرة - مصر
تليفون : ٢٥٢٧٥٢٢٠ / ٥ / ٦ / ٧ (٢٠٢)
فاكس : ٢٥٢٧٥٢٢٤ (٢٠٢)

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To

The Quality of Industrial Products Testing Laboratories.
General Organization for Export & Import Control (GOEIC).
(Alexandria Port Branch – Gate18 - Alexandria Governorate - Egypt)

Schedule No.: 041905B 1st Accreditation Date: January 23, 2020 Issue No. (02): February 28, 2024 Revision No. (-): Valid to: February 27, 2028

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Water Pumps	Performance Test of Guarantee point test for pump Flow rate: 30 L/min to 370L/min Head up to 150 m	ES 4104:2018 ISO/IEC9906-2012 Pump Test Instrument Model: PTE 100S- 30/300 LPM S.N.: 980042 LINE 1 (1") Flow Meter S.N: 024007966 LINE 2 (2") Flow Meter S.N: 024007391 Pressure Sensor S.N:4031913989

Almery

Farid El Denawi